

July ____, 2011

[Customer Name
Customer Address
Customer Address]

RE: THERMO SCIENTIFIC HM355S-3 USER SAFETY ADVISORY NOTICE

Dear ...:

Microm is dedicated to providing customers with the highest quality instruments and we consider user safety to be a significant priority. Microm has identified a need to enhance the safety instructions and the description of the safety devices in the user manual. To this end, we are modifying the user manual of the HM355S-3 as follows:

1. A modification to the chapter regarding safe work practices for the instrument as well as additional instructions on proper brake usage has been made more specific.
2. A new Quick-Reference-Card is being provided which summarizes the enhancements made to the user manual.

To minimize inconvenience to your organization, the updated user manuals are enclosed with this letter on CD. Please print the updated user manual so that it may be readily available for reference near your instrument. If you cannot print the updated user manual or you would like a hard copy, please contact Technical Support at the number below. In addition, a new Quick-Reference-Card is enclosed with this letter. Please instruct your laboratory personnel to complete the Confirmation of Receipt Form and return it to us by September 30, 2011 as well as review the materials and incorporate safeguards into current work processes.

If you or any of your customers have additional questions or need further information or assistance, please contact one of our Technical Support representatives at:

Attn: Factory service support Technical Support
Subject: HM355S User Safety Advisory Notice
Ph: +49 (0) 6227 836-25499 (08:00 – 16:30 / Germany)
Fax: +49 (0) 6227 836 - 111
Email: fss.dxd.dewal@thermofisher.com

Sincerely,



Christof Rosswag
Managing Director
Microm International GmbH,
a part of Thermo Fisher Scientific

CONFIRMATION OF RECEIPT FORM

Product Information Notification

So that we can complete our records, please return this form to us by **September 30, 2011**.

FAX TO: Houssam Raad, HM355S-3 User Manual Advisory Notice
FAX: +49 (0)6227 836 - 111

I hereby confirm that I have received the new User Manual and Quick-Reference-Card of the HM355S-3 rotary microtome.

Name: _____

Title: _____

HM355S-3 Serial #: _____

Tel. number: () _____

Firm name: _____

Address: _____

City/state: _____