



**MEDICAL DEVICE RECALL LISTING NOVEMBER 2024**

<b>Date Received</b>	<b>Reference No.</b>	<b>Recall Type</b>	<b>Product Name</b>	<b>Product Registration</b>	<b>Recall Class</b>	<b>Reason of Recall</b>	<b>Recalling Establishment</b>	<b>Establishment License</b>
<b>11/11/2024</b>	MDA/Recall/P0346-47411217-2024	Voluntary Recall	HYDROPHOBIC ACRYLIC INTRAOCULAR LENS	GC7143222-113629	Class III	A02: Manufacturing, Packaging or Shipping Problem	MY IOL SDN BHD	MDA-6395-K124
<b>11/11/2024</b>	MDA/Recall/P0347-99297066-2024	Voluntary Recall	BD BBL™ SENSI-DISC™ SUSCEPTIBILITY TEST DISCS	IVDB6578123-148932	Class II	A23: Use of Device Problem	BECTON DICKINSON SDN BHD	MDA-5083-W123
<b>26/11/2024</b>	MDA/Recall/P0349-25322329-2024	Voluntary Recall	3M UNITEK TRANSBOND PLUS SELF ETCHING PRIMER	GB71798433017	Class III	A02: Manufacturing, Packaging or Shipping Problem	3M MALAYSIA SDN BHD	MDA-5752-WP124

\* The information contained in the Medical Device Authority Recall database is released under Regulation 7(8) and Regulation 8 (5) of Malaysia’s Medical Device Regulations 2019.