



Reference: 2021-002M

9th August 2021

URGENT - FIELD SAFETY NOTICE

To all users of Olympus Endoscope Reprocessor OER-AW

Re: Endoscope Label on Front Door of Olympus Endoscope Reprocessor OER-AW

Dear Customer,

Olympus is implementing a Field Corrective Action for the device listed below. The affected model and serial number is listed below.

Model Number	Model Description	Serial Number
OER-AW	Endoscope Reprocessor	2933077, 2933100, 2933101, 2933113, 2033164, 2033295, 2033297, 2033298, 2033299, 2033300, 2033305, 2033306, 2033307,

Olympus has become aware of an issue that requires your attention. This notice pertains to the consistency of the categories for Endoscope setup displayed on the “ENDOSCOPE COMBINATIONS FOR DUAL SCOPE REPROCESSING” label on the front door of OER-AW endoscope reprocessor.

On the label affixed on the front door of the affected OER-AW, the TJF-Q290V endoscope model is not correctly categorized because the TJF-Q290V must not be combined with any other endoscope and must be reprocessed individually.

The categorization of the TJF-Q290V and the fact this model must be reprocessed individually is correctly indicated on the warning label at the top lid of the OER-AW as well as the devices’ Instructions for Use (IFU).

Olympus has not received any complaint associated with the incorrect categorization of the TJF-Q290V endoscope on the label affixed on the front door of the OER-AW. However, it is possible that a customer could become confused about the proper endoscope setup when reprocessing a TJF-Q290V scope with an OER-AW.

To prevent this confusion, Olympus will replace the label currently on the front door of your OER-AW with the correct label.

**ENDOSCOPE COMBINATIONS FOR DUAL SCOPE REPROCESSING
FOR PINK COLORED RETAINING RACK (MAJ-1970)**

WARNING ENDOSCOPES LISTED IN GROUP 1 (GASTROINTESTINAL ENDOSCOPES OR BRONCHOSCOPES) SHOULD NOT BE REPROCESSED IN COMBINATION WITH ENDOSCOPES LISTED IN GROUP 2 (SURGICAL ENDOSCOPES). THESE COMBINATIONS MAY LIMIT THE EFFECTIVENESS OF CLEANING AND DISINFECTION.

- ANY TWO ENDOSCOPES LISTED IN THE GROUP 1 CAN BE REPROCESSED SIMULTANEOUSLY.
- ANY TWO ENDOSCOPES LISTED IN THE GROUP 2 CAN BE REPROCESSED SIMULTANEOUSLY.
- THIS CHART APPLIED ONLY FOR THE OER-AW WITH PINK COLORED RETAINING RACK (MAJ-1970). DO NOT USE THIS CHART FOR THE OER-AW WITH WHITE COLORED RETAINING RACK (GL830600).

NOTE THESE ENDOSCOPES MAY NOT BE AVAILABLE IN SOME AREAS.

<GROUP 1> (GASTROINTESTINAL ENDOSCOPES OR BRONCHOSCOPES)				<GROUP 2> (SURGICAL ENDOSCOPES)			
MODEL	FIBERSCOPE	VIDEOSCOPE		MODEL	FIBERSCOPE	VIDEOSCOPE	
GIF	1T130	140	H190	N260	Q240	XP160	XQ260
	E3	160	H260	P140	Q240X	XP170N	XTQ160
	K20	1T140	H260Z	PG260	Q240Z	XP180N	
	P30	1T240	H290	PV70	Q260	XP190N	
	PG20	1TH190	H290EC	Q140	Q260J	XP240	
	Q40	1TQ160	H290Z	Q145	SP240	XP260	
	XP20	FQ260Z	HQ190	Q150	V	XP260N	
	XPE3		H170	HQ290	Q160	V	XP260NS
	XQ40		H180	KQ240	Q160Z	V70	XP290N
			H180J	LV1	Q165	XK240	XQ140
			H185	N180	Q180	XP150N	XQ240
	GF		Q240				
	JF	1T140 20 TE3	140F	240	260V		
	TJF	90	140F	160R	Q180V		
M20		140R	160VF	Q290V			
		145	160VR				
		150	240				
		160F	260V				
CF	40L/I	140L/I/S	H170L/I	H260DL/I	Q140L/I	Q165L/I	V70L/I
	E3L/I	1T140L/I	H180AL/I	H290EC	Q145L/I	Q180AL/I	VL/I
	EL/I	200S	H180DL/I	H290L/I	Q150L/I	Q240AL/I	
		240AL/I	H185L/I	HG190L/I	Q160AL/I	Q240L/I	
		240DL/I	H190L/I	HG290L/I	Q160DL/I	Q240ZL/I	
		240L/I	H260AL/I	HG290ZL/I	Q160L/I/S	Q260AL/I	
		FH260AZL/I	H260AZL/I	LV1L/I	Q180ZL/I	Q260DL/I	
PCF	20	140L/I	H190TL/I	PH190L/I	Q260JL/I		
	E3L/I	160AL/I	H290DL/I	PG260L/I	S		
		240L/I	H290L/I	Q180CAL/I			
		H180AL/I	H290TL/I	Q240Z			
		H190DL/I	H290ZL/I	Q260AL/I			
		H190L/I	F240AL/I	Q260CAZ			
CSF	3 4	V60					
CHF	BF30						
BF	1T130	PE	160	1TQ290	MP180F	Q180	
	1T140	PE2	1T150	240	MP290F	Q180-AC	
	1T160	TE	1T160	260	P150	Q190	
	3C40	TE2	1T180	3C160	P160	Q290	
	40	XP40	1T240	6C240	P180	XP160F	
	MP60	XP60	1T240R	6C260	P190	XP190	
	N20	XT30	1T260	F260	P240	XP260F	
	P30	XT40	1TH190	H190	P260F	XP290	
	P40	MP190F	1TQ170	H290	P290	XT160	
	P60		1TQ180	MP160F	Q170	XT190	

THIS CHART IS BASED UPON INFORMATION AVAILABLE AS OF MMM DD, YYYY. PLEASE CONTACT OLYMPUS REGARDING ENDOSCOPES INTRODUCED AFTER MMM DD, YYYY OR REGARDING ANY QUESTIONS ABOUT ENDOSCOPE COMBINATIONS LISTED IN THIS CHART.

*1 Steam sterilization is preferred reprocessing method for this model, but it can also be high-level disinfected in the OER-AW.

Figure 1. Incorrect “ENDOSCOPE COMBINATIONS FOR DUAL SCOPE REPROCESSING” label

Action(s) to be taken by end user:

Our records indicate that you have purchased the affected product(s) and we request you to take the following action(s):

1. Please read through carefully the enclosed information and share this notice to any healthcare professional from your organization as appropriate.
2. Please inspect your inventory of OER-AW device to identify any of the specified serial number listed in above table.
3. Complete and return the ‘Response Form’ to Olympus.
4. Upon receiving your ‘Response Form’, an Olympus representative will contact you to replace the label at the front door of your OER-AW device.

OLYMPUS (MALAYSIA) SDN. BHD. (200101010901)

Lot No. B-6-2, Level 6, The Ascent, Paradigm, No. 1, Jalan SS7/26A, Kelana Jaya, 47301 Petaling Jaya, Selangor
Tel: (603) 7886 9188 Fax : (603) 7887 2833



Olympus regrets any inconvenience caused and fully appreciates your prompt cooperation in addressing this situation. If you have any questions or concerns, please do not hesitate to contact us for additional information.

Contact for enquiries

Regulatory Affairs Department

Email : mes-ra.oml@olympus.com

Tel : (603) 7886 9188

Fax : (603) 7887 2833

The **Medical Device Authority** has been informed of this notice.

Yours sincerely,

A handwritten signature in black ink, consisting of the Japanese characters '坂口 有史' (Sakaguchi Yuji).

.....
Yuji Sakaguchi

Managing Director

Olympus (Malaysia) Sdn. Bhd.



Response Form

Please send the complete and signed Response Form to Regulatory Affairs Department at:

To : Olympus (Malaysia) Sdn. Bhd, Regulatory Affairs
 Fax/Email : (603) 7887 2833 / mes-ra.oml@olympus.com
 From : _____ [Facility Name] Contact no.: _____
 Date : _____
 Ref : 2021-002M

URGENT - FIELD SAFETY NOTICE

Re: Endoscope Label on Front Door of Olympus Endoscope Reprocessor Model OER-AW

I acknowledge receipt of the Field Safety Notice (“FSN”) referenced above. I understand that I need to undertake the action(s) listed in the FSN.

Check the applicable boxes below:

- I DO NOT have affected devices remaining. All have been used or discarded.
- I DO have the affected devices, which I will adhere to the FSN.

Serial Number	

Name: _____

Designation: _____

.....
Signature & Company Stamp

.....
Date