

Initial Report Form

Field Safety Corrective Action

Medical Devices Vigilance System
(MEDDEV 2.12/1 rev 6)

1. Administrative information

Destination

Name of national competent authority (NCA)

Address of national competent authority

Date of this report

1 April 2011

Reference number assigned by the manufacturer

PRA 2001-004

Incidence reference number and name of the co-ordinating national competent authority (if applicable)

Identify to what other national competent authorities this report was also sent

Belgium, France, Hungary, Italy, Slovenia, UK, Switzerland

2 Information on submitter of the report

Status of submitter

- Manufacturer
- Authorized representative within EEA
- Others (identify the role):

3 Manufacturer information

DePuy Spine
325 Paramount Drive
Raynham, MA 02767
USA

Manufacturer's contact person

Andrew Topoulos

Address

As above

Postal code

As above

City

As above

Phone

+1 508 828-2790

Fax

N/A

E-mail

N/A

Country

As above

4 Authorized representative information

Name of the authorized representative

DePuy International Ltd
St Anthony's Road
Leeds LS11 8DT
England

The authorized representative's contact person

Paul Arnott

Address

As above

Postal code

As above

City

Leeds

Phone

+44 (0) 7771 971 930

Fax

+44 113 3876087

E-mail parnott@its.jnj.com	Country England
5 National contact point information	
National contact point name As above	
Name of the contact person As above	
Address As above	
Postal code As above	City As above
Phone As above	Fax As above
E-mail As above	Country As above
6 Medical device information	
Class	
<input type="checkbox"/> AIMD Active implants <input type="checkbox"/> MDD Class III <input type="checkbox"/> MDD Class IIb <input checked="" type="checkbox"/> MDD Class IIa <input type="checkbox"/> MDD Class I	
<input type="checkbox"/> IVD Annex II List A <input type="checkbox"/> IVD Annex II List B <input type="checkbox"/> IVD Devices for self-testing <input type="checkbox"/> IVD General	
Nomenclature system (preferable GMDN) GMDN Code = 12727	
Nomenclature code	
Nomenclature text Intended Use of the Device: The Confidence Needles are supplied sterile. They are used to place bone cement percutaneously at a desired location. They come in a variety of sizes (length and diameter) and configurations (Bevel tip, Diamond tip). The cement extrudes from the end of the cannulated needle with the exception of the side hole needle which extrudes cement from a portal on the side of the tube near the distal end. The needle is placed under fluoroscopy prior to cement delivery. The needle is also sometimes used in MIS cases to place the guide wire that is used to assist in the placement of the pedicle screw.	
Commercial name/brand name/make DePuy Spine Confidence® Introducer Needle	
Catalogue Numbers / Description: 283903611 Confidence 11g x 6-inch Introducer Needle, Diamond Tip 283904413 Confidence 13g x 4-inch Introducer Needle, Side Hole Lot Number: HLPB4G & HLPB4F	
Serial number(s) and/or lot/batch number(s) As above	
Software version number (if applicable) N/A	
Manufacturing date/expiry date (if applicable) Not applicable	
Accessories/associated device (if applicable) Not applicable	
Notified body (NB) ID- number BSi – 0086	

7 Description of FSCA
<p>Background information and reason for the FSCA</p> <p>The reason for this Field Safety Corrective Action and Field Safety Notice is that a lot which is properly labeled (11g x 6 inch) contains a 13g x 4-inch side hole needle with the sterile package. As the needle contained within the package is two inches shorter in length it could result in an inability to get the shorter needle to the desired location. If no additional 6-inch needles were on hand the procedure may have to be aborted. The second lot was in production at the same time as the side fire needle which is also part of this field safety corrective action.</p>
<p>Description and justification of the action (corrective/preventive)</p> <p>As above</p>
<p>Advice on actions to be taken by the distributor and the user</p> <p>Recalled products identified above must not be used. Notify your DePuy Sales Representative or Account Manager as soon as possible.</p>
<p>Attached please find</p> <p><input checked="" type="checkbox"/> Field Safety Notice (FSN) in English</p>
<p>Time schedule for the implementation of the different actions</p> <p>Immediate</p>
<p>These countries within the EEA and Switzerland are affected by this FSCA</p> <p>Within EEA and Switzerland:</p> <p>Belgium, France, Hungary, Italy, Slovenia, UK, Switzerland</p> <p>Candidate Countries:</p>
<p>These countries outside the EEA and Switzerland are affected by this FSCA</p> <p>Australia, Malaysia, USA, South Africa, UAE</p>
8 Comments

I affirm that the information given above is correct to the best of my knowledge.



Signature

Name

City

Date

Paul Arnott

Leeds, UK

April 2011

Submission of this report does not, in itself, represent a conclusion by the manufacturer and/or authorized representative or the national competent authority that the content of this report is complete or accurate, that the medical device(s) listed failed in any manner and/or that the medical device(s) caused or contributed to the alleged death or deterioration in the state of the health of any person.