



**MEDICAL DEVICE RECALL LISTING JULY 2024**

Date Received	Reference No.	Recall Type	Product Name	Product Registration	Recall Class	Reason of Recall	Recalling Establishment	Establishment License
01/07/2024	MDA/Recall/P0302-75727245-2024	Voluntary Recall	CAPIO SLIM SUTURE CAPTURING DEVICE	GB71551564718	Class II	A24: Adverse Event Without Identified Device or Use Problem	BOSTON SCIENTIFIC (MALAYSIA) SDN BHD	MDA-5810-WD124
03/07/2024	MDA/Recall/P0303-51703664-2024	Voluntary Recall	ALINITY C PROTEINS	IVDB5170819-34873	Class II	A08: Calibration Problem	ABBOTT LABORATORIES (MALAYSIA) SDN. BHD.	MDA-5104-W123
05/07/2024	MDA/Recall/P0304-42114480-2024	Voluntary Recall	VASCUGRAFT® NEO	GC37117108716	Class III	A21: Labelling, Instructions for Use or Training Problem	B. BRAUN MEDICAL INDUSTRIES SDN. BHD.	MDA-4250-W123
09/07/2024	MDA/Recall/P0305-15896132-2024	Voluntary Recall	DSD EDGE ENDOSCOPE REPROCESSOR	GC16706518418	Class II	A21: Labelling, Instructions for Use or Training Problem	STERIS MALAYSIA SDN. BHD.	MDA-5302-WDP123
09/07/2024	MDA/Recall/P0309-17658684-2024	Voluntary Recall	NIM-NEURO 3.0 SYSTEM AND ACCESSORIES	GB45802845818	Class II	A23: Use of Device Problem	MEDTRONIC MALAYSIA SDN BHD	MDA-4793-WDP123
18/07/2024	MDA/Recall/P0311-30582816-2024	Voluntary Recall	MCGRATH MAC LARYNGOSCOPE	GB10147922-107333	Class III	A27: Appropriate Term/Code Not Available	MEDTRONIC MALAYSIA SDN BHD	MDA-4793-WDP123
23/07/2024	MDA/Recall/P0312-10509171-2024	Voluntary Recall	SUSCEPTIBILITY AND IDENTIFICATION TEST FOR MYCOBACTERIA	IVDC23724200218	Class II	A23: Use of Device Problem	BECTON DICKINSON SDN BHD	MDA-5083-W123

\* The information contained in the Medical Device Authority Recall database is released under Regulation 7(8) and Regulation 8 (5) of Malaysia’s Medical Device Regulations 2019