

URGENT: VOLUNTARY MEDICAL DEVICE RECALL	
Description	Voluntary recall of specific lots of DAILIES TOTAL1® contact lenses
Product Reference	DAILIES TOTAL1 MULTIFOCAL (DELEFILCON A) SOFT CONTACT LENSES

August 8, 2024

Dear Valued Customer,

Alcon has initiated a Voluntary Medical Device Recall for specific lots of DAILIES TOTAL1® MFOCAL contact lenses. An isolated quality issue was identified with a material supplied by a third-party vendor that was used in the production of contact lenses at a single manufacturing site. As a result, the affected lenses may not meet Alcon quality and/or performance standards for the entirety of their labeled shelf life.

The contact lenses from the identified lots are not expected to cause an increased risk to the wearer. As with any contact lens, there is a remote likelihood that wearing a lens from the affected lots may lead to temporary issues such as irritation, redness, or blurred vision. These symptoms typically resolve after lens removal and/or replacement (as directed in the patient leaflet).

According to our records, affected contact lens lots have been dispensed to you. Please discontinue use of the product and contact the outlet where you received the contact lenses for further information on next steps.

Sincerely,