



**MEDICAL DEVICE RECALL LISTING DECEMBER 2024**

<b>Date Received</b>	<b>Reference No.</b>	<b>Recall Type</b>	<b>Product Name</b>	<b>Product Registration</b>	<b>Recall Class</b>	<b>Reason of Recall</b>	<b>Recalling Establishment</b>	<b>Establishment License</b>
<b>03/12/2024</b>	MDA/Recall/P0348-41448449-2024	Voluntary Recall	CENTRIFUGE	GA84458804218	Class III	A05: Mechanical Problem	INTERSCIENCE SDN BHD	MDA-5690-W124
<b>05/12/2024</b>	MDA/Recall/P0350-82624946-2024	Voluntary Recall	STEALTHSTATION S7 TREATMENT GUIDANCE SYSTEM (SPINAL)	GD2205223-131802	Class II	A23: Use of Device Problem	MEDTRONIC MALAYSIA SDN BHD	MDA-4793-WDP123
<b>10/12/2024</b>	MDA/Recall/P0353-70976527-2024	Voluntary Recall	SELDINGER CHEST DRAINAGE KITS	GB2435122-96118	Class III	A02: Manufacturing, Packaging or Shipping Problem	SOMNOTECH (M) SDN BHD	MDA-4866-WDP123
<b>18/12/2024</b>	MDA/Recall/P0354-13401169-2024	Voluntary Recall	MECHANICAL LITHOTRIPTOR	GA65102107116	Class II	A21: Labelling, Instructions for Use or Training Problem	OLYMPUS (MALAYSIA) SDN. BHD.	MDA-6344-WDP124

\* The information contained in the Medical Device Authority Recall database is released under Regulation 7(8) and Regulation 8 (5) of Malaysia’s Medical Device Regulations 2019.