

Comorde Medical (M) Sdn Bhd Announces Voluntary Nationwide Recall of **Adult 1.2µm Lipid Filter**. Due to Manufacturing Defect.

- 1 Date of Report: **6 November 2025**
- 2 MDA Recall reference no: **MDA/Recall/P0460-20270414-2025**
- 3 MDA reporting ref no: **MDA/MPR/P25842-40674648-2025**

4 Product Information:

Product Name	Identifier	Affected Lot	QC Pass Date
Adult 1.2µm Lipid Filter	CM-LF-1.2µm-15-15	202503FPS0025	2 July 2025
Adult 1.2µm Lipid Filter	CM-LF-1.2µm-15-15	202503FPS0025	3 July 2025
Adult 1.2µm Lipid Filter	CM-LF-1.2µm-15-15	202503FPS0025	9 July 2025

5 Product Distribution Information

Hospital Kuala Lumpur	ACCUMED TECHNOLOGY SDN BHD
3000	1000

Comcorde Medical (M) Sdn Bhd is voluntarily recalling all lots of all **Adult 1.2µm Lipid Filter** of the above-mentioned identifier and lot number due to manufacturing defect where there might be a chance of tube being detached from the filter during use. It may have the potential to raise adverse effects such as blood loss towards the patient.

The level of recall is to the user, that is, regional hospital pharmacies and related departments, and physician's office practices. This mandatory recall is to ensure patient safety and product quality

To return the unused product, please contact your local sales representatives

Alternatively, you may complete the enclosed Verification Form and e-mail to feedback@comcorde.com.my to imitate the return of the device

For audit purposes, please return the form to state weather you still have inventory of the product even if you choose not to return it

Please be assured that replacement product will be available by end November 2025 that will address this issue. In the meantime, we apologize for any inconvenience this may cause and appreciate your attention to this matter

Signature 
Name Wong Kah Yoong
Designation Quality Management Representative
Date 6 Nov 2025

Product Recall Reply Form

(1) Date: _____

(2) From (Person in charge) _____

(3) Name of Establishment _____

(4) Phone no _____

(5) Email _____

I/we, **Person in charge(2)** of **Establishment(3)** has read and acknowledge the Product Recall Report on page1

Identifier	Lot No	Batch/QC Pass	Quantity (pcs)
Adult 1.2µm Lipid Filter	202503FPS0025	2 July 2025	
Adult 1.2µm Lipid Filter	202503FPS0025	3 July 2025	
Adult 1.2µm Lipid Filter	202503FPS0025	9 July 2025	

I/we wish to return the following remaining stock that is in my/our **establishment (3)**

With this reply, any difference in quantity from point 5 in page 1 was either used or disposed and that I/we no longer hold any other stock other than the one mentioned in the table above.

Signature _____

Name _____

Date _____